



BSI Standards Publication

**High frequency surgical equipment and
high frequency surgical accessories -
Operation and maintenance**

National foreword

This Published Document is the UK implementation of IEC TR 61289:2019.

The UK participation in its preparation was entrusted to Technical Committee CH/62/4, Electromedical equipment.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2019
Published by BSI Standards Limited 2019

ISBN 978 0 580 99925 3

ICS 11.040.01

Compliance with a British Standard cannot confer immunity from legal obligations.

This Published Document was published under the authority of the Standards Policy and Strategy Committee on 31 May 2019.

Amendments/corrigenda issued since publication

Date	Text affected
------	---------------



IEC TR 61289

Edition 2.0 2019-05

TECHNICAL REPORT



**High frequency surgical equipment and high frequency surgical accessories –
Operation and maintenance**

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

ICS 11.040.01

ISBN 978-2-8322-6885-8

Warning! Make sure that you obtained this publication from an authorized distributor.

CONTENTS

FOREWORD	4
INTRODUCTION	6
1 Scope	7
2 Normative references	7
3 Terms and definitions	7
4 General information regarding HF SURGICAL EQUIPMENT	12
5 Recommended practices before use	13
5.1 Inspection of HF SURGICAL EQUIPMENT and HF SURGICAL ACCESSORIES before use	13
5.2 Preparation	13
5.3 Handling of NEUTRAL ELECTRODES, HF SURGICAL ACCESSORIES, cables and connections	14
5.4 PATIENTS with active implants (active implantable medical devices)	14
5.5 Simultaneous use of two items of HF SURGICAL EQUIPMENT	15
5.6 Instructions for use	15
6 Recommended practices during use	15
7 Recommended practices after use	16
8 Nature of HAZARDS	16
8.1 General	16
8.2 HF SURGICAL EQUIPMENT related HAZARDS	16
8.2.1 Incompatible combinations	16
8.2.2 Electromagnetic compatibility	16
8.2.3 Misconnection of ACTIVE ACCESSORIES	16
8.2.4 Specialty HF SURGICAL EQUIPMENT	17
8.3 ACTIVE ACCESSORY related HAZARDS	17
8.3.1 Incompatible combinations	17
8.3.2 Environment of use	17
8.3.3 Misuse	18
8.4 OPERATOR-related HAZARDS	18
8.4.1 OPERATOR not reading or following the instructions for use	18
8.4.2 OPERATOR selecting inappropriate power or mode settings	18
8.4.3 OPERATOR using an HF SURGICAL ACCESSORY in an inappropriate manner	18
8.5 NEUTRAL ELECTRODE related HAZARDS	19
8.5.1 General	19
8.5.2 Inadequate contact area of a NEUTRAL ELECTRODE	19
8.5.3 Inappropriate application	19
8.5.4 Surgical procedures utilizing high currents or long duty cycles	19
9 Safety provisions of, and symbols on, HF SURGICAL EQUIPMENT	22
9.1 General	22
9.2 Colours of indicator lights	22
9.3 Markings on HF SURGICAL ELECTRICAL EQUIPMENT	22
9.4 Protection against electric shock and burns	24
9.4.1 Method of protection	24
9.4.2 Degree of protection	24
9.5 HF SURGICAL EQUIPMENT not properly marked	25
9.6 Monitoring the effectiveness of the NEUTRAL ELECTRODE	25

9.7	Output indicators	25
10	Accompanying documents	26
11	Preventive maintenance	26
Annex A (informative) Concerns regarding high current surgical procedures or high duty cycles.....		27
Bibliography.....		28
Figure 1 – Symbol used on Class II equipment		24
Figure 2 – Symbol used with an earth referenced PATIENT circuit		25
Figure 3 – Symbol used with an HF ISOLATED PATIENT CIRCUIT		25
Table 1 – Colours and significance of indicator lights according to IEC 60601-2-2		22
Table 2 – Symbols used on HF SURGICAL EQUIPMENT		23

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**HIGH FREQUENCY SURGICAL EQUIPMENT AND HIGH FREQUENCY
SURGICAL ACCESSORIES – OPERATION AND MAINTENANCE**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

The main task of IEC technical committees is to prepare International Standards. However, a technical committee may propose the publication of a technical report when it has collected data of a different kind from that which is normally published as an International Standard, for example "state of the art".

IEC 61289, which is a technical report, has been prepared by sub-committee 62D: Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice.

This second edition cancels and replaces the first edition published in 2011. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) general adaption to IEC 60601-2-2:2017;
- b) refinement and additions to the defined terms;

- c) separation of HF SURGICAL EQUIPMENT and HF SURGICAL ACCESSORIES;
- d) consideration of the HIGH CURRENT MODE;
- e) update of symbols.

The text of this technical report is based on the following documents:

Enquiry draft	Report on voting
62D/1652/DTR	62D/1662A/RVDTR

Full information on the voting for the approval of this technical report can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this document, the following print types are used:

– TERMS DEFINED IN CLAUSE 3: SMALL CAPITALS.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

A bilingual version of this publication may be issued at a later date.

IMPORTANT – The “colour inside” logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this publication using a colour printer.

INTRODUCTION

This document gives guidelines to personnel in charge of operation of equipment covered by IEC 60601-2-2:2017 to enable them to attain the best conditions of safety for their PATIENTS and themselves.

HIGH FREQUENCY SURGICAL EQUIPMENT AND HIGH FREQUENCY SURGICAL ACCESSORIES – OPERATION AND MAINTENANCE

1 Scope

This document contains guidelines for medical and nursing personnel regarding the safe and effective operation of HIGH FREQUENCY SURGICAL EQUIPMENT and HIGH FREQUENCY SURGICAL ACCESSORIES (also referred to as HF SURGICAL EQUIPMENT in this document). It is also of use to scientific/technical staff who have responsibility for the maintenance of this equipment.

The application guidelines in this document deal with the safe operation of HIGH FREQUENCY SURGICAL EQUIPMENT constructed according to the safety requirements of IEC 60601-1 [1]¹ and IEC 60601-2-2 [4].

Not all existing HIGH FREQUENCY SURGICAL EQUIPMENT meets the minimum requirements of current international standards, however, the guidelines in this document is still helpful in utilizing these devices.

2 Normative references

There are no normative references in this document.

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

3.1

ACCESSORY

additional part for use with equipment in order to:

- achieve the INTENDED USE,
- adapt it to some special use,
- facilitate its use,
- enhance its performance, or
- enable its functions to be integrated with those of other equipment

[SOURCE: IEC 60601-1:2005, 3.3]

3.2

ACTIVE ACCESSORY

HF SURGICAL ACCESSORY intended for manipulation by the OPERATOR to produce an effect by electrical conduction adjacent to the ACTIVE ELECTRODE at the intended site on the PATIENT,

¹ Numbers in square brackets refer to the Bibliography.