



BSI Standards Publication

# **Ambient air — Determination of the particle number concentration of atmospheric aerosol**

**National foreword**

This Published Document is the UK implementation of CEN/TS 16976:2016.

The UK participation in its preparation was entrusted to Technical Committee EH/2/3, Ambient atmospheres.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2016.  
Published by BSI Standards Limited 2016

ISBN 978 0 580 90926 9

ICS 13.040.20

**Compliance with a British Standard cannot confer immunity from legal obligations.**

This Published Document was published under the authority of the Standards Policy and Strategy Committee on 31 August 2016.

**Amendments/corrigenda issued since publication**

Date	Text affected
------	---------------

---

TECHNICAL SPECIFICATION  
SPÉCIFICATION TECHNIQUE  
TECHNISCHE SPEZIFIKATION

**CEN/TS 16976**

August 2016

ICS 13.040.20

English Version

**Ambient air - Determination of the particle number  
concentration of atmospheric aerosol**

Air ambiant - Détermination de la concentration en  
nombre de particules de l'aérosol atmosphérique

Außenluft - Bestimmung der  
Partikelanzahlkonzentration des atmosphärischen  
Aerosols

This Technical Specification (CEN/TS) was approved by CEN on 26 June 2016 for provisional application.

The period of validity of this CEN/TS is limited initially to three years. After two years the members of CEN will be requested to submit their comments, particularly on the question whether the CEN/TS can be converted into a European Standard.

CEN members are required to announce the existence of this CEN/TS in the same way as for an EN and to make the CEN/TS available promptly at national level in an appropriate form. It is permissible to keep conflicting national standards in force (in parallel to the CEN/TS) until the final decision about the possible conversion of the CEN/TS into an EN is reached.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels**

## Contents

Page

European Foreword.....	5
Introduction .....	6
1 Scope .....	7
2 Normative references .....	7
3 Terms and definitions .....	7
4 Atmospheric aerosol .....	8
5 Description of the method .....	9
5.1 Sampling and conditioning .....	9
5.1.1 Sampling .....	9
5.1.2 Drying .....	10
5.1.3 Dilution .....	11
5.2 Determination of the number concentration with a CPC .....	11
5.2.1 Condensation growth .....	11
5.2.2 Optical detection .....	13
6 CPC performance criteria and test procedures .....	14
6.1 General .....	14
6.2 General requirements of the CPC .....	14
6.3 Test conditions .....	15
6.4 Performance characteristics and criteria .....	15
6.5 Test procedures .....	16
6.5.1 Inlet flow rate accuracy .....	16
6.5.2 Number concentration measurement range .....	16
6.5.3 Number concentration detection limit .....	16
6.5.4 Linearity and slope of response .....	16
6.5.5 Detection efficiency curve at low particle size .....	17
6.5.6 Upper particle size detection limit .....	17
6.5.7 Zero count rate .....	17
6.5.8 Response time .....	17
6.5.9 Dependence of flow rate on supply voltage .....	17
6.5.10 Accuracy of temperature and pressure sensor calibration .....	18
6.5.11 Effect of failure of mains voltage .....	18
7 Performance criteria and test procedures for the sampling and conditioning system .....	18
7.1 General requirements .....	18
7.2 Performance characteristics and criteria .....	18
7.3 Diffusion losses .....	19
7.4 Relative humidity .....	19
7.5 Dilution factor .....	19
7.6 Primary sampling flow .....	19
8 Measurement procedure .....	20
8.1 Measurement planning .....	20
8.2 Environmental operating conditions .....	20
8.3 Initial installation .....	20
8.4 Initial checks on site .....	20

8.5	Data processing and reporting .....	21
9	Quality control, quality assurance and measurement uncertainty .....	21
9.1	General.....	21
9.2	Frequency of calibrations, checks and maintenance.....	21
9.2.1	General.....	21
9.2.2	Maintenance of CPC.....	22
9.2.3	Calibration of CPC plateau region and linearity .....	22
9.2.4	Determination of CPC low size cut-off.....	22
9.2.5	CPC zero check .....	22
9.2.6	Number concentration check .....	23
9.2.7	CPC flow rate calibration.....	23
9.2.8	Temperature and pressure sensor calibration.....	23
9.2.9	CPC internal diagnostics.....	23
9.2.10	Sample system maintenance .....	23
9.2.11	Relative humidity sensor .....	23
9.2.12	Dilution factor (where applicable) .....	23
9.2.13	Leak check.....	23
9.3	Measurement uncertainty .....	24
9.3.1	General.....	24
9.3.2	CPC plateau detection efficiency .....	24
9.3.3	CPC detection efficiency drift.....	24
9.3.4	Flow determination .....	25
9.3.5	Correction to standard temperature and pressure.....	25
9.3.6	Sampling losses due to diffusion to walls.....	25
9.3.7	Dilution factor (where applicable) .....	25
9.3.8	Calculation of overall uncertainty .....	25
	Annex A (normative) Determination of diffusion losses in sampling lines .....	27
	Annex B (informative) Example of the calculation of diffusion losses in a sampling system .....	29
B.1	Description of the sampling system.....	29
B.2	Air properties and diffusion coefficient .....	30
B.3	Losses in the primary sampling tube.....	30
B.4	Losses in the secondary sampling tube and the dryer .....	31
B.5	Overall sampling losses .....	31
	Annex C (informative) Data reporting.....	32
C.1	Motivation .....	32
C.2	Level 0 (annotated raw data) .....	32
C.3	Level 1 (data processed to final physical property, potential corrections applied, original temporal resolution) .....	33
C.4	Level 2 (hourly averages, including measures of variability) .....	33
C.5	GAW WDCA Condensation Particle Counter Level 0 (raw data) file format example (system without sample dilution): .....	35
C.6	Line-by-line explanations: .....	35
	Annex D (informative) Uncertainty calculation (example) .....	47
D.1	General.....	47
D.2	CPC plateau detection efficiency .....	47

<b>D.3</b>	<b>CPC detection efficiency drift.....</b>	<b>47</b>
<b>D.4</b>	<b>Flow determination.....</b>	<b>47</b>
<b>D.5</b>	<b>Correction to standard temperature and pressure .....</b>	<b>47</b>
<b>D.6</b>	<b>Sampling losses due to diffusion to walls.....</b>	<b>47</b>
<b>D.7</b>	<b>Dilution factor (where applicable) .....</b>	<b>48</b>
<b>D.8</b>	<b>Calculation of overall uncertainty.....</b>	<b>48</b>
	<b>Annex E (informative) Atmospheric aerosols in Europe .....</b>	<b>49</b>
<b>E.1</b>	<b>General.....</b>	<b>49</b>
<b>E.2</b>	<b>Mean concentrations .....</b>	<b>49</b>
<b>E.3</b>	<b>Examples of measurements.....</b>	<b>50</b>
	<b>Annex F (informative) Dilution systems .....</b>	<b>53</b>
<b>F.1</b>	<b>Background .....</b>	<b>53</b>
<b>F.2</b>	<b>Criteria for dilution systems .....</b>	<b>53</b>
<b>F.3</b>	<b>Operation principles of dilution systems.....</b>	<b>53</b>
<b>F.3.1</b>	<b>General.....</b>	<b>53</b>
<b>F.3.2</b>	<b>Dilution systems with partial flow filtration.....</b>	<b>54</b>
<b>F.3.3</b>	<b>Dilution systems with external clean air supply .....</b>	<b>54</b>
	<b>Annex G (informative) Laminar flow.....</b>	<b>56</b>

## **European Foreword**

This document (CEN/TS 16976:2016) has been prepared by Technical Committee CEN/TC 264 “Air quality”, the secretariat of which is held by DIN.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

According to the CEN/CENELEC Internal Regulations, the national standards organisations of the following countries are bound to announce this Technical Specification: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

## Introduction

There is a growing awareness of the significance of aerosol particles with diameters of  $D < 1 \mu\text{m}$  for human health as well as for their climatic impact. To assess air quality, it appears necessary to supplement gravimetrically determined mass concentrations such as  $\text{PM}_{10}$  or  $\text{PM}_{2.5}$  with a measurement of the particle number concentration. Since ultrafine particles with diameters of  $D < 0,1 \mu\text{m}$  make an almost insignificant contribution to the mass of atmospheric aerosol particles, they can best be detected with counting measuring methods of sufficient sensitivity.

As particle measurement instrumentation allows determining either the particle number concentration or the particle number size distribution two Technical Specifications will be established:

- one dealing with the determination of the single parameter number concentration (a measure of “total” number concentration),
- one dealing with the determination of number concentration within a limited number of size ranges.

Clauses 5 and 6 contain general information about the method and the expected properties of the aerosol to be measured.

Clause 7 sets out the performance criteria for CPCs. Specifically, these are the relevant performance characteristics of CPC instruments (without any sampling system), the respective criteria that shall be met, and a description of how the tests shall be carried out. In general these tests are expected to be carried out by test houses or CPC manufacturers rather than users, and could form the basis for type testing of CPCs in future.

Clause 8 sets out the performance criteria and test procedures for the sampling and conditioning system (e.g. dilution). These may be applied by manufacturers of sampling systems, test houses or users (network operators).

Clause 9 sets out requirements for the installation, initial checks and calibrations, and operation of a CPC and sampling system at a monitoring site, including routine maintenance, data processing (including use of QA/QC data) and reporting. In general these will be the responsibility of users (network operators), though calibrations requiring test aerosols shall only be carried out by suitably qualified laboratories.

Clause 10 sets out Quality Assurance and Quality Control procedures, i.e. the ongoing checks and calibrations that are required on the CPC and sampling system during operation at a monitoring site. It is expected that these will be the responsibility of users (network operators), though calibrations requiring test aerosols shall only be carried out by suitably qualified laboratories. The main sources of measurement uncertainty are described.

## 1 Scope

This Technical Specification describes a standard method for determining the particle number concentration in ambient air in a range up to about  $10^7 \text{ cm}^{-3}$  for averaging times equal to or larger than 1 min. The standard method is based on a Condensation Particle Counter (CPC) operated in the counting mode and an appropriate dilution system for concentrations exceeding the counting mode range. It also defines the performance characteristics and the minimum requirements of the instruments to be used. The lower and upper sizes considered within this document are 7 nm and a few micrometres, respectively. This document describes sampling, operation, data processing and QA/QC procedures including calibration parameters.

## 2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 27891, *Aerosol particle number concentration — Calibration of condensation particle counters*

## 3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

### 3.1

#### **actual flow rate**

volumetric flow rate of an individual instrument, measured at its inlet under the actual air conditions

### 3.2

#### **aerosol**

a multi-phase system of solid and/or liquid particles suspended in a gas, ranging in particle size from  $0,001 \mu\text{m}$  to  $100 \mu\text{m}$

### 3.3

#### **calculation flow rate**

flow rate which directly relates count rate and particle number concentration

Note 1 to entry: This flow rate is used for instrument internal calculation of the particle number concentration. It depends on the instrument type and may be nominal, factory-certified or actual inlet flow rate. It may also include a calibration factor unless the total inlet flow is analysed.

### 3.4

#### **coincidence error**

error that occurs with counting measuring methods when two or more particles are counted simultaneously as a single particle

Note 1 to entry: Coincidence error is related to particle number concentration, flow velocity through the sensing zone and size of sensing zone.