



BSI Standards Publication

Clinical laboratory testing — Criteria for acceptable lots of dehydrated Mueller-Hinton agar and broth for antimicrobial susceptibility testing

National foreword

This Published Document is the UK implementation of ISO/TS 16782:2016.

The UK participation in its preparation was entrusted to Technical Committee CH/212, IVDs.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2016.

Published by BSI Standards Limited 2016

ISBN 978 0 580 95546 4

ICS 11.100.20

Compliance with a British Standard cannot confer immunity from legal obligations.

This Published Document was published under the authority of the Standards Policy and Strategy Committee on 30 November 2016.

Amendments/corrigenda issued since publication

Date	Text affected
------	---------------

TECHNICAL SPECIFICATION

**ISO/TS
16782**

First edition
2016-10-15

Clinical laboratory testing — Criteria for acceptable lots of dehydrated Mueller-Hinton agar and broth for antimicrobial susceptibility testing

*Détermination de la sensibilité aux antibiotiques — Critères
d'acceptabilité pour les lots d'agar déshydraté et de bouillon Mueller-
Hinton pour déterminer la sensibilité aux antibiotiques*



Reference number
ISO/TS 16782:2016(E)

© ISO 2016



COPYRIGHT PROTECTED DOCUMENT

© ISO 2016, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative reference	1
3 Terms and definitions	1
4 Requirements for Mueller-Hinton broth	3
4.1 Components of Mueller-Hinton broth	3
4.2 Physical and chemical characteristics	3
4.2.1 Dehydrated powder or granules	3
4.2.2 Prepared broth medium	4
4.2.3 Cation supplementation and content for MHB	4
4.2.4 Other medium components	4
4.2.5 Specific adjustments required by the manufacturer	4
4.3 Manufacturers protocol for testing production lots of dehydrated Mueller-Hinton broth	5
4.4 Interpreting the results	5
4.5 Evaluating the results	6
5 Requirements for Muller-Hinton agar	6
5.1 Components of Mueller-Hinton agar	6
5.2 Physical and chemical characteristics	6
5.2.1 Dehydrated powder or granules	6
5.2.2 Prepared agar medium	7
5.2.3 Cation supplementation and content for MHA	7
5.2.4 Other medium components	7
5.2.5 Specific adjustments required by the manufacturer	7
5.3 Manufacturer's protocol for testing production lots of dehydrated Mueller-Hinton agar	8
5.4 Interpreting the results	8
5.5 Evaluating the results	10
6 Testing new antimicrobial agents with production lots of dehydrated Mueller-Hinton broth or agar	11
Annex A (informative) Mueller-Hinton medium	12
Annex B (informative) Preparing control cultures	14
Annex C (informative) Suggested data sheet for testing of production lots	16
Annex D (informative) Label statement	19
Bibliography	20

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing Technical Specifications is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary information](#)

The committee responsible for this document is ISO/TC 212, *Clinical laboratory testing and in vitro diagnostic test systems*.

Introduction

Historically, although various media have been recommended for susceptibility testing, Mueller-Hinton broth (MHB) has been selected as the medium for the reference broth microdilution minimum inhibitory concentration (MIC) method (ISO 20776-1) and Mueller-Hinton agar (MHA) is most widely used for disc diffusion testing of rapidly growing bacteria.

Mueller-Hinton medium provides satisfactory growth of most non-fastidious pathogens, acceptable batch-to-batch reproducibility, low sulfonamide, trimethoprim, and tetracycline inhibitors and a large amount of data has been collected from antimicrobial susceptibility tests with this medium over several decades.

This International Standard is the result of an effort to establish a standard description and protocol by which manufacturers of dehydrated Mueller-Hinton agar (dMHA) and broth (dMHB) may determine its acceptable performance characteristics.

The results of testing conform to defined quality control limit ranges for each combination of antimicrobial agent and quality control strains. Each production lot is tested at least against these combinations of antimicrobial agents and quality control strains.

This Technical Specification has been developed in part based upon two Clinical and Laboratory Standards Institute (CLSI) documents, CLSI M6-A2^[1] (protocols for evaluating dehydrated Mueller-Hinton agar) and CLSI M32-P^[2] (evaluation of lots of dehydrated Mueller-Hinton broth for antimicrobial susceptibility testing) with permission. Upon publication of ISO 16782, CLSI documents M6-A2^[1] and M32-P^[2] will no longer be available. Manufacturers can follow ISO 16782 to assess the performance characteristics of their production lots of dMHA and dMHB.

Clinical laboratory testing — Criteria for acceptable lots of dehydrated Mueller-Hinton agar and broth for antimicrobial susceptibility testing

1 Scope

This Technical Specification provides a standard description of the physical properties of dehydrated Mueller-Hinton broth (dMHB) and Mueller-Hinton agar (dMHA) and performance criteria by which manufacturers can assess the performance characteristics of their production lots of dMHA and dMHB. Production lots of broth or agar can then be utilized by all users, including *in vitro* susceptibility testing device manufacturers, as the test medium for performance of antimicrobial susceptibility testing.

This Technical Specification does not address supplements (e.g. blood or blood products) that are added to the medium to support growth of fastidious bacteria[3][4][5][6]. Those additives are provided after the dehydrated medium is prepared in its liquid state as a final product and fall outside of the scope of this Technical Specification. Although dMHA can be used for determination of MICs using the agar dilution method[4][6] or the gradient diffusion method, this Technical Specification only includes performance testing of dMHA using disc diffusion methodology as described by the Clinical and Laboratory Standards Institute (CLSI)[5] and European Committee on Antimicrobial Susceptibility Testing (EUCAST)[3].

2 Normative reference

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 20776-1:2006, *Clinical laboratory testing and in vitro diagnostic test systems — Susceptibility testing of infectious agents and evaluation of performance of antimicrobial susceptibility test devices — Part 1: Reference method for testing the in vitro activity of antimicrobial agents against rapidly growing aerobic bacteria involved in infectious diseases*

CLSI M100, *Performance Standards for Antimicrobial Susceptibility Testing; Informational Supplement*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

antimicrobial agent

substance of biological, semi-synthetic or synthetic origin that inhibits the growth of or kills bacteria and is thus of potential use in the treatment of infections

Note 1 to entry: Disinfectants, antiseptics and preservatives are not included in this definition.

[SOURCE: ISO 20776-1:2006, 2.1]

3.2

antimicrobial disc

small paper disc containing known amounts of antimicrobial agents used for *in vitro* susceptibility testing